



217 Osseo Ave. No. • St. Cloud, Minnesota 56303
(320) 253-1640 • (800) 599-BUGS (2847) • Fax (320) 253-6250
www.microbiologics.com

Statistical Analysis Certificate

Company:

Product Description: EZ-CFU™ Microorganisms

Microorganism Name: Staphylococcus aureus subsp. aureus

Reference #: ATCC 6538

Catalog #: 0485C

Lot #: 485000

Expiration Date: 2010-03

Mean Assay Value: 4.0×10^3 CFU per mL

Standard Deviation: 9.2×10^2

Coefficient of Variation: 23%

99% Confidence Interval of 3.8×10^3 to 4.2×10^3 CFU

95% Confidence Interval of 3.8×10^3 to 4.1×10^3 CFU

Method used to determine Mean Assay Value: Spiral Biotech Test Method

Medium Employed: Standard Methods Agar

Incubation Time and Temp: 24 hrs at 34-38 degrees C

Certificate Prepared By:

Date: